

ITALY(PEAKS) Technical consultation (to-do) - Request #1017

[LIST #12] FLEXTRAC Bulk loader sample overflow (FLEXTRAC 0000 00 00 000000)

07/14/2026 09:41 AM - Kang Min

Status:	in progress (00)	Start date:	07/14/2026
Priority:	00 (medium)	Due date:	
Assignee:	Che Emma	% Done:	0%
Category:		Estimated time:	0.00 hour
Target version:		Spent time:	0.00 hour
Description FLEXTRAC Bulk loader sample overflow While the TSS has a medium-small collection tank, the Flextrac bulk loader has a very small collection tank (see images). If a test tube leaks, the reservoir is already full. If another test tube leaks or falls into the reservoir, liquid will overflow, contaminating the surrounding area or potentially damaging the sensors. In addition, absorbent material must be placed in this tank to absorb any liquid that may fall. TSS00 00 0000 0000 00 00, Flextrac 0000 0000 00 00 00000 0000 00000(0000 00). 0000 00000 0000 00000 00 00 00 0000. 00000 0000 00000 0000 0000000 0000 0000000 0000 00, 0000 00 0000 000000000 0000 00000 0 00000. 00, 0 0000 00000 0000 00 0000 0000 0000 0000.			

History

- #1 - 07/14/2026 09:41 AM - Kang Min
- [PEAKS]
- the section is tilted in order to collect the liquid toward the front area, where a sensor detects the presence of material and alerts the customer EN can provide absorbent materials.
- 03/03: the sensor can not detect liquid at present they are working to improve. it is recommended to have a absorbing material in place.
- 17/03: It is necessary to check where the spilled liquid will end up, then we will discuss about tray capability and sensor.
- 29/04: let us know when you have checked
- #4 - 07/14/2026 09:45 AM - Kang Min
- Status changed from Receipt (00) to in progress (00)
- #5 - 07/14/2026 09:47 AM - Kang Min
- Tracker changed from Conference to Request
- #6 - 07/14/2026 09:49 AM - Kang Min
- [ENERGIUM]
- Internal meeting completed. Will revert after internal confirmation.
- 23/04: The sensor can not detect liquid
- #7 - 07/14/2026 09:53 AM - Kang Min
- Assignee changed from Kang Min to Che Emma
- #8 - 07/14/2026 11:58 AM - KIM Dooki
- Project changed from Technical consultation (to-do) to ITALY(PEAKS) Technical consultation (to-do)